

## CALIBRATION CERTIFICATE

Frigga certifies that the product(s) mentioned on the following page(s) have been thoroughly calibrated under the observation of ISO9001:2015 certified quality assurance system and meet performance accuracy specifications over the stated ranges.

### 1. Reference instruments

Model	No. of Instrument	Accuracy	Measure Range
Fluke 51-II	56110274WS	±0.05	-200~100°C

Note: All reference instruments used in validation are periodically calibrated by CNAS (China National Accreditation Service) accredited laboratories, CNAS is full member of the International Laboratory Accreditation Cooperation, so this Calibration Certificate is also accepted by all signatories to the ILAC MRA.

### 2. Environment Conditions

Ambient Temperature: 20~25°C; Ambient Humidity: (50~60)%RH; Others: /

### 3. Validation Result

Validation points	Tolerance	Specification	Result
-196°C	0.4	±1.0°C	Passed
-80°C	0.2	±1.0°C	Passed
-20°C	0.2	±0.5°C	Passed
0°C	0.0	±0.5°C	Passed

### 4. Product information

Product Name	Device ID	Quantity
T7	T2814001	1

Calibration expires 12 months from the date of sale.

Certificate by: *Carol*

Approved by: *Daisy*

Date: 2026-01-13

